

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 20 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>Scott E. Gardner</u>	ID# <u>6222</u> Date <u>14 May 10</u>
A Agency <u>Directorate of Emergency Services</u> Phone # <u>384-3888</u>	
Instrument Location <u>Bldg #656, A St, Ft Richardson, AK 99505</u>	
B Alco S/N <u>301542</u> Target Value <u>.082</u> High Pressure <u>875 psi</u>	
Alco Test Values	<u>.083</u> 1 <sup>st</sup> Alco
	<u>.084</u> 2 <sup>nd</sup> Alco
Signature <u>Scott E. Gardner</u>	<u>CB</u> <u>5/24/10</u>
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 17<sup>th</sup> day of June, 2010.

Carolyn M. Noland  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130142

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142

MAY 14, 2018

OPERATOR'S NAME:

GARDNER/SCOTT/

OPERATOR'S NUMBER: 6222

SUBJECT'S LAST NAME:

N/A

SUBJECT'S FIRST NAME/MI :

N/A

O.L. #: UOC

DEPT/AGENCY: FRM

CASE/REPORT: VERIFICATION OF CAL

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: 1301542

05/24/10

### — BREATH ANALYSIS —

.002 ADJUSTED FOR 29.64 in

ALCO TARGET .001 17:16

BLANK TEST .000 17:17

INTERNAL STANDARD VERIFIED 17:17

ALCO TV 29.64 in .003 17:17

BLANK TEST .000 17:18

SUBJECT SAMPLE .000 17:19

BLANK TEST .000 17:19

ALCO TV 29.65 in .004 17:20

BLANK TEST .000 17:21

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142

MAY 14, 2018

TIME 17:14

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 45c

BAROMETER: 29.64 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~